INSTRUCTION MANUAL

DIGIVAC 200 and 801W Series Wide Range Thermocouple Vacuum Gauges and Controllers

Model 200 (benchtop) Model 200P and 200P/2c (panel mounting) Model 801W and 801W/2c (Varian pattern panel mounting)


Model Dvg6 and Dvg4 (Hastings pattern panel mounting) Optional Controls, analog output, recorders and Rs232

> <u>Ranges</u> 760 torr to .001 torr 760 torr to .1 millitorr 1000 mbar to .001 mbar

> > Digivac Company (formerly Fairfield Digital) 83 Bridgewaters Rd. Oceanport, NJ, 07757 WWW.digivac.com (732) 571-2300 (732) 571-4697 FAX

DIGIVAC 200 and 801W Series Wide Range Thermocouple Vacuum Gauges and Controllers


Model 200 (benchtop)



Model 200P and 200P/2c (panel mounting)



Model 801W and 801W/2c (Varian or Hastings pattern, panel mounting)



Model dVt-6 and dVt-4 (Hastings pattern panel mounting) Optional Controls, analog output, recorders, and Rs232

> Ranges 760 torr to .001 torr 760 torr to .0001 torr 1000 mbar to .001 mbar

Digivac Model 200 / 200P / 801W Wide Range Thermocouple Gauges Benchtop / Panel Mount with optional controls and Rs232

Thank you for purchasing a DIGIVAC[™] 200 or 801W series Digital Vacuum Gauge. These gauges read from atmospheric pressure to full vacuum (0 microns) using a rugged, low cost, thermocouple vacuum gauge tube.

The Model 200 is bench mounting while the Model 200P is designed for panel mounting in customers' equipment.

The Model 200P requires a panel cutout 1.65" inches high by 3.65" inches wide. The gauge is approximately 4.5" deep behind the panel. If panel depth is a problem, consider the 801W gauge which is only 1.5" deep and mounts neatly on panel doors.

The Model 801W is electrically identical to the other gauges, but is designed to be a bolt-in replacement for Varian 801 and other panel mounting analog vacuum gauges. The Digivac 810-2 has the exact mounting dimensions of the Varian 810-2 gauge. These gauges may also be ordered with a 4 bolt mounting pattern to directly replace Hastings gauges, while still reading to atmosphere. To install these gauges, remove the front panel and install the gauge from the rear.

Both gauges are also available with one or two control outputs. Either Model can also be ordered with a serial port or an analog recorder output.

LIMITATIONS

These Digivac gauges use thermocouple technology. As such they have a very wide range (six decades!) and reasonable accuracy.

Because they work on heat dissipation, they are dependent on the thermal conductivity of the residual gasses.

Other Digivac gauges use transducer and capacitance manometer technology. These gauges have a range of 0 to 760 torr and a resolution of 1 and 0.1 torr and an accuracy of +/- 0.3% of full scale. At the low end they have a repeatability of +/- 1 torr. These gauges are not dependent on gas composition. The tradeoff then is resolution and range against accuracy.

MOUNTING INSTRUCTIONS FOR 801(W) AND DVG-6 SERIES GAUGES

The Digivac 801, 801w, Dvg-4, and Dvg-6 gauges mount in the same panel cutouts as the respective Varian, Hastings, and Televac gauges they are replacing.

For new installations, proceed as follows:

- Remove the front panel of the gauge, leaving the aluminum spacers on the gauge printed circuit board.
- Make a 2.66" circular mounting hole in the panel at the desired location
- Make 3 of 4 holes, 5/32" diameter (clearance holes for \$4 screws) on a 1.60 radius from the center of the mounting hole.
 (The front panel that comes with the gauge can be used as a template.)

Reassemble the gauge, with the plastic panel outside, and the gauge behind.

Digivac Dvg4 and Dvg6 gauges mount in a similar manner, except for having more mounting holes.

GAUGE TUBE INSTALLATION

The tube must be mounted in a stem down position within 10 degrees of vertical. Its connection is one-eighth inch male pipe thread. Thread tape will make sealing easier. Be careful not to twist the cable.

The tube should be installed in a clean part of the vacuum system. It is best not to mount it directly on a vacuum pump suction as the tube life may be shortened by backstreaming of pump oil vapors. Inserting a loop of tubing or an optical baffle is often helpful in protecting the tube from backstreaming.

After energizing, the unit will initialize and operate. It reaches best thermocouple accuracy after a few minutes of warm-up. For greater accuracy Digivac offers a combination transducer/thermocouple gauge which offers 1% accuracy in Torr ranges. We also offer 1/% accurate capacitance manometer gauges.

The gauges have a switching power supply that can be connected to either 115 or 220 VAC, 50/60 hz.

The relay outputs of the Model 200P should be connected at this time if they are to be used. The connection positions are marked on the terminal strip. Note: not all 200P units have control outputs. All terminal strips have markings. The outputs are SPDT, 250 VAC, 7 amps. If the contacts are switching highly inductive loads, varistors or other transient protective components may be required. The Normally open contacts close when the vacuum setpoint is reached.

At atmosphere, the unit will read 760. A visible decimal point indicates that the unit is reading in Torr. As pressures are decreased, the unit's readings will shift to 99.0, 9.99, .999, and finally to .001 One micron is equal to one millitorr.

The reading of the gauge will be influenced by the composition of the gasses remaining. All wide range and Piriani vacuum gauges are subject to this phenomenon. For applications in which excellent accuracy is required in the low torr ranges, such as the filling of neon and laser tubes, we offer a transducer-based guage which gives precise readings regardless of gas composition. Transducer gauges are not able to read in the micron ranges.

Representative readings are as follows:

, 1997 - 1997

Torr	millitorr or microns
760.	760,000
550.	550,000
120.	120,000
99.0	99,000
12.3	12,300
10.0	10,000
9.90	9,900
1.23	1,230
.999	999
.555	555
.123	123
.055	55
.001	1

Note: The Digivac Model 100W, preferred by many in the refrigeration industry reads directly in microns.

TROUBLESHOOTING

If the gauge reads 760 torr and does not respond to vacuum, it is likely that the thermocouple voltage is not getting to the gauge. The cause could be:

Bad thermocouple Bad cable connection

replace gauge tube On Model 200P, check cable wires at terminal strip On Model 200, check red, black, green and white wires on 9 pin dsub connector On all units, check soldered connections On octal connector

If the gauge reads 0:

Bad thermocouple	replace gauge tube
Bad cable connections	See above


RECALIBRATION

Addinate state

If the tube is replaced, the unit must be recalibrated. It is best that this be done at Digivac. If field calibration is done, the zero and span adjustments are available on two potentiometers which are accessible by removing the front panel of the instrument. The zero is on the right and the span is on the left. Make slight adjustments and wait a few seconds for the effects of the adjustments to be seen. On the 801 series gauges, the zero adjustment is located on the front panel.

Atmosphere

- · Energize the unit at atmospheric pressure. Remove the front panel.
- Make sure that the gauge tube is in a vertical orientation, as it is position sensitive.
- Connect a Voltmeter to the two pads available from the right side of the printed circuit board, these are called Zero and Atm .420
- Adjust the potentiometer labeled "Atm" (leftmost) until the voltage above equals .420
- This gives an approximate adjustment. Now expose the tube to a pressure of 600 millimeters of mercury and adjust for 600 on the display.



Model 200 with cover removed

5

<u>Zero</u>

Operate the unit at the lowest attainable pressure.

- 1 Micron with a diffusion pump,
- 10 to 15 Microns from a belt-driven two-stage mechanical pump,
- 35 Microns from a direct-drive, two-stage mechanical pump,
- · 500 Microns from a single-stage mechanical pump,
- 50 torr (50,000 microns) or higher for dry mechanical and diaphragm pumps).

Connect Voltmeter leads to "Com" and "Vac" pins.

The voltage should be approximately .161 Volts dc.


Make slight adjustments, if necessary, so the low end vacuum reading is correct.

If the above readings cannot be attained, either the instrument, the gauge tube, or the cable is defective. Gauge tubes often become contaminated with oil. Cleaning the gauge tube with a strong solvent, such as a carburetor cleaner can often improve or restore its calibration.

Factory recalibration, performed with a diffusion pump, is available for \$65.00 (cal with data is extra)

At right is an 801/810 series unit with the front panel removed. The atmosphere pot is at the top and the zero is at the bottom in the center.

The various test points are at the top of the board and are identified with writing. For user recalibrations, it is only necessary to adjust the zero pot. It can be reached through a hole in the front panel, so disassembly is not necessary.




CONTROL SETPOINTS

To adjust the control setpoints of the Model 200P, carefully remove the display lens. This is accomplished by gently prying the bottom of the lens with a small screwdriver (available from Radio Shack). On the left of the instrument are two potentiometers and a three position jumper strip. The jumper strip has a shorting bar attached to it. When this bar is removed, the unit displays indicated vacuum. When placed so as to short out the left two pins, the unit will display the #1 setpoint. When placed on the right, it displays #2. Place the jumper for the desired setpoint, and adjust the potentiometer.

On the Model 800 series gauges with controls, the setpoints are adjusted by pressing front panel mounted buttons and turning the adjacent potentiometers. When the button is pressed, the setpoint shows up on the display.

Model 200P with red lens removed



Rs-232 OUTPUT

On units equipped with Rs232, the current value on the screen is sent out the Rs232 port approximately once per second.

The parameters are: 9600 baud, 8 data bits, 1 stop bit, no parity.

No prompting or handshaking by the user is required.

The data is presented as ascii text for the numbers (with decimal points as required), followed by a carriage return and line feed.

This data may be observed using Windows Terminal, Hyperterm, or any similar terminal program.

Data may be saved using Terminal's "Transfers / Receive Text File" feature. The saved data files may be read with Microsoft Excel or any text editor.


Microsoft Excel, or any spreadsheet program can also plot the data.

Digivac also offers a monitoring program that show the data on the screen and opens and names a log file in response to a single click by the user.

A companion program, Digiplot plots the data. Wide range data is plotted logarithmically. The current file is selected automatically, or older files may be opened.

The Rs232 connection is make on the terminal strip of units such as the Model 200P and the terminals are identified as Gnd and Tx.

Other units have 9 pin dsubminiature connectors, and the data appears on pin 2, and the ground is on pin 5.



DIGIVAC AND DIGIPLOT

For more convenient logging and plotting, the Digivac Company offers Digivac and DigiPlot software. Custom Visual Basic applications are also available.


- The Digivac program gives a visual presentation of the data which is coming out of the vacuum gauge.
- Update rate is approximately once per second.
- When "Log Start" is pressed a logging file with a backup is opened. The rate of data logging (eg, once per minute) can be set by the user.
- Digiplot, the companion plotting program, opens the current logging file with a single mouse click.
- Logarithmic axes clearly show vacuum from atmosphere to 10 ^ 3 or 10 ^ 9 (ultra wide range version).
- One more click prints out the plots in color.
- A window at the bottom of the plot screen show the data file name and enables the user to scroll through values.

П

REAR PANEL CONNECTIONS


The Model 200P has a terminal strip and all of its connections are marked. The following is an explanation of these markings.

- Negative connection from 5 volt power supply.
- + This is + 5 volts output from the power supply.
- Rec Recorder (Linear) output on units so equipped.
- Gnd Ground for outputs. (connect to pin 5 of a 9 pin D-sub for Rs-232)
- Txd RS-232 transmit (connect to pin 2 of a 9 pin D-sub
- No Normally open Relay contact
- Nc Normally closed relay contact
- Com Common Relay contact.
- X No Connection
- G Green wire to gauge tube
- R Red wire to gauge tube
- B Black wire to gauge tube
- W White wire to gauge tube
- (usually signal -) (usually signal +)
- (usually exitation +)
- (usually exitation -)



DIGIVAC RECORDER

Digivac offers a special vacuum recorder which displays from Atmosphere to hard vacuum. Both the plot and the numbers are displayed. This recorder is often used by rebuilders to document the performance of vacuum pumps but have many other uses. They are priced at under \$400.00 and are available from stock. Contact Digivac Company for further information.



CUSTOM FEATURES AND PROGRAMMING

Because these gauges are microprocessor controlled, it is possible to provide custom control features at surprisingly reasonable prices. The software is "burned" into the microprocessors, which are socketed and easily replaced by a user. Contact the Digivac Company to discuss your needs.

If the gauge is to be used in a <u>neon sign</u> processing facility, it must be installed with a stopcock or a solenoid valve to protect it from exposure to the system and possible damage during high voltage bombarding. Battery operated gauges are also offered and are particularly suited to neon applications.

For repair or recalibration, return gauges to:

The Digivac Company 83 Bridgewaters Road Oceanport, NJ, 07757 (732) 571 2300 (732) 571 4697 Fax

e-mail: direct from our website www.digivac.com

The Digivac Company manufactures a complete line of vacuum gauges and computers. Contact us or your distributor if you wish for further information.

See www.digivac.com for our latest offerings